

Abstract Submitted
for the DNP13 Meeting of
The American Physical Society

SIPT—An Ultrasensitive Mass Spectrometer for Rare Isotopes¹

SAMUEL J. NOVARIO, GEORG BOLLEN, DAVID L. LINCOLN, ADRIAN A. VALVERDE, National Superconducting Cyclotron Laboratory/Michigan State University, RYAN RINGLE, STEFAN SCHWARZ, National Superconducting Cyclotron Laboratory, MATTHEW REDSHAW, National Superconducting Cyclotron Laboratory/Central Michigan University, LEBIT TEAM — Over the last few decades, advances in radioactive beam facilities like the Coupled Cyclotron Facility at the National Superconducting Cyclotron Laboratory (NSCL) at Michigan State University (MSU) have made short-lived, rare-isotope beams available for study in various science areas, and new facilities, like the Facility for Rare Isotope Beams (FRIB) under construction at MSU, will provide even more exotic rare isotopes. The determination of the masses of these rare isotopes is of utmost importance since it provides a direct measurement of the binding energy of the nucleons in the atomic nucleus. For this purpose we are currently developing a dedicated Single-Ion Penning Trap (SIPT) mass spectrometer at NSCL to handle the specific challenges posed by rare isotopes. These challenges, which include short half-lives and extremely low production rates, are dealt with by employing the narrowband FT-ICR detection method under cryogenic conditions. Used in concert with the 9.4-T time-of-flight mass spectrometer, the 7-T SIPT system will ensure that the LEBIT mass measurement program at MSU will make optimal use of the wide range of rare isotope beams provided by the future FRIB facility.

¹Funding from NSF grant PHY-1126282 is acknowledged

Samuel J. Novario
National Superconducting Cyclotron Laboratory/Michigan State University

Date submitted: 01 Jul 2013

Electronic form version 1.4