

Abstract Submitted
for the DPP05 Meeting of
The American Physical Society

Development of an Active Plasma Control System for Pegasus¹

M.W. BONGARD, S.P. BURKE, R.J. FONCK, B.T. LEWICKI, University of Wisconsin-Madison — The Phase II Pegasus ST experiment includes fully programmable power supplies for all magnet coils. These will be integrated with a digital feedback plasma control system (PCS), based on the PCS in use on DIII-D, to provide active feedback control of the plasma evolution. The initial goal is to control $I_p(t)$, $R(t)$, and $Z(t)$. The feedback cycle consists of: 1) sampling magnetic diagnostics, 2) applying a response matrix derived from equilibrium reconstructions, 3) accounting for induced vessel currents and power supply frequency responses, and 4) adjusting the current demand, all on a timescale that is fast compared to the shot duration of ~ 25 ms. Data is sampled via a 500 kHz, 96-channel CPCI digitizer from DTACQ Solutions, Ltd. The power supply control signals are generated by 16 analog waveform generators, with the option to utilize 32 digital I/O lines in the future. The PCS digitizer is controlled via a system of Linux-based computers that perform requisite computation-intensive tasks and interface to the existing LabVIEW control codes via a TCP/IP network link.

¹Work supported by U.S. D.O.E. Grant DE-FG02-96ER54375.

Gregory Garstka
University of Wisconsin-Madison

Date submitted: 22 Jul 2005

Electronic form version 1.4