

Abstract Submitted
for the DPP07 Meeting of
The American Physical Society

New Optics for the Soft X-ray Diagnostic on DIII-D¹ M.J. LANCOT, Columbia U., E.M. HOLLMANN, UCSD, R.K. FISHER, S. PIDCOE, D.A. TAUSSIG, General Atomics — The optics for the soft x-ray poloidal array on the DIII-D tokamak have been upgraded to include a new set of 64 photodiodes and an adjustable filter wheel. The wheel includes three titanium-coated diamond filters for measuring soft x-ray emission and two unfiltered settings for fast bolometry. We present the specifics of the design, including filter transmission functions, and techniques used to reduce pickup noise from nearby coils. Recent results from the revamped system will be analyzed with a focus toward categorizing MHD instabilities observed in DIII-D plasmas.

¹Supported by the US DOE under DE-FG02-89ER53297, DE-FG02-04ER54758, DE-FC02-04ER54698, and a Fusion Energy Science Fellowship.

M.J. Lancot
Columbia U.

Date submitted: 23 Jul 2007

Electronic form version 1.4