

Abstract Submitted
for the DPP07 Meeting of
The American Physical Society

***H α* power loss in SSPX** T.L. STEWART, C.A. ROMERO-TALAMAS, H.S. MCLEAN, D.L. CORRELL, Lawrence Livermore National Laboratory, Livermore, CA 94550, SSPX TEAM — An absolute calibration has been performed on each of the seven chords of the *H α* diagnostic [Z. Wang, G.A. Wurden, C.W Barnes, et al., Rev. Sci. Instrum. 72, 1059 (2001)] at SSPX. Simple models are used to estimate the total power lost to *H α* radiation throughout experimental shots. Using these models, high energy shots ($T_e > 500$ eV) are compared to low energy shots.

C. A. Romero-Talamas
Lawrence Livermore National Laboratory, Livermore, CA 94550

Date submitted: 24 Jul 2007

Electronic form version 1.4