

Abstract Submitted
for the MAR06 Meeting of
The American Physical Society

Sagnac type fiber interferometer for magneto-optic Kerr effect measurement at cryogenic temperatures¹ JING XIA, Stanford University, PETER BEYERSDORF, San Jose State University, MARTIN FEJER, AHARON KAPITULNIK, Stanford University — We describe a Sagnac type magneto-optical interferometer operating at 1550 nm wavelength in which the fast and slow axis of a single 10-meter long Polarization-Maintaining fiber were used as the Sagnac loop. The last 2 meters of this PM fiber were fed into a cryogenic probe to measure Kerr rotation. This zero-area-Sagnac-loop design is virtually immune to temperature fluctuations and mechanical bending of the fiber, and can work at much lower temperature (< 1 K) compared to apparatus with optical windows. Most important, no perturbing AC magnetic field is needed. Up to now, we have achieved a Kerr rotation sensitivity of $2.5 \times 10^{-7} \text{ rad}/\sqrt{Hz}$ down to liquid Helium temperature, with $4 \mu W$ of optical power at the detector. Drifts of this apparatus were observed to be less than $0.5 \times 10^{-7} \text{ rad}/\text{Hour}$. We studied ferromagnetic transition and magnetic domains of thin (3 to 30 nm) SrRuO₃ films by cooling them through Tc in zero fields (< 5 mG) and measuring polar Kerr rotations.

¹supported by CPN, NSF grant PHY-0425897

Jing Xia
Stanford University

Date submitted: 12 Jan 2006

Electronic form version 1.4