

Abstract Submitted
for the MAR07 Meeting of
The American Physical Society

A low-field SQUID MRI system: design, performance and characterization.¹ BYEONG HO EOM, KONSTANTIN PENANEN, INSEOB HAHN, Jet Propulsion Laboratory/California Institute of Technology — A low-field MRI system with SQUID detector and gradiometer sense coil readout is being developed for clinical imaging. The system is compact and low-weight and operates in minimally shielded environment. An in-vivo image of human hand with sensitive field of view of 3cm, in-plane resolution of 1mm and 4mm slice width can be acquired in ~ 5 minutes. We discuss the system design, optimization, noise characteristics and imaging performance of the apparatus.

¹This work is supported by NASA

Byeong Ho Eom
Jet Propulsion Laboratory/California Institute of Technology

Date submitted: 17 Nov 2006

Electronic form version 1.4