

Abstract Submitted
for the MAR16 Meeting of
The American Physical Society

Amylin Detection with a Miniature Optical-Fiber Based Sensor

ZHAOWEN LIU, MATSKO ANN, ADAM HUGHES, MARK REEVES, George Washington University — We present results of a biosensor based on shifts in the localized surface plasmon resonance of gold nanoparticles self-assembled on the end of an optical fiber. This system allows for detection of protein expression in low sensing volumes and for scanning in cell cultures and tissue samples. Positive and negative controls were done using biotin/avidin and the BSA/Anti-BSA system. These demonstrate that detection is specific and sensitive to nanomolar levels. Sensing of amylin, an important protein for pancreatic function, was performed with polyclonal and monoclonal antibodies. The measured data demonstrates the difference in sensitivity to the two types of antibodies, and titration experiments establish the sensitivity of the sensor. Further experiments demonstrate that the sensor can be regenerated and then reused.

Mark Reeves
George Washington Univ

Date submitted: 06 Nov 2015

Electronic form version 1.4