

Abstract Submitted  
for the APR07 Meeting of  
The American Physical Society

**Magnetic Field Reduction in Photomultipliers with the use of Bucking Coil** DALGIS MESA, JOERG REINHOLD, Florida International University — Three segmented aerogel Cerenkov detectors are employed in the Jefferson Lab High Resolution Kaon Spectrometer (HKS). Cerenkov light is detected with 5” PMTs. At the detector location a stray field from the magnet of  $\sim 5$  gauss significantly reduces the photoelectron collection efficiency of the PMTs, despite the use of magnetic shielding. Therefore, we propose to compensate the stray field with bucking coils. In a test setup, a Helmholtz Coil was used to generate external magnetic fields up to 15 gauss. In order to counteract this  $\mathbf{B}$  field 20 turns of bucking coil (gauge: 12 AWG) were mounted around the cathode sides of the PMT. A recovery of roughly 87.5% was determined for an external  $\mathbf{B}$  field of 5 Gauss. The experimental procedure and results of both this and new studies will be presented in detail.

Dalgis Mesa  
Florida International University

Date submitted: 16 Jan 2007

Electronic form version 1.4