

Abstract Submitted
for the DPP07 Meeting of
The American Physical Society

Validating the simulation of the effects of secondary neutrals on the MSE gas-filled-torus calibration¹ WILLIAM SCHUMAKER, Lawrence Technological University, HOWARD YUH, FRED LEVINTON, Nova Photonics, Inc., STEVE SCOTT, Princeton Plasma Physics Laboratory — A common procedure of injecting a neutral beam into a gas-filled torus with known magnetic fields in vacuum is used to calibrate motional Stark effect (MSE) diagnostics on many toroidal magnetic devices. A cause of anomalies encountered in this calibration has been explained as a consequence of secondary neutrals from the ionized beam. Under certain conditions, these re-neutrals emit H-alpha spectra that have the proper Doppler shift to pass through the MSE filters yet have a different polarization than those from the primary beam neutrals, thus contaminating the measured Stark electric field angle. Existing IDL code has been adapted to simulate the gas-filled-torus calibration of MSE in Alcator C-Mod, NSTX, and TFTR vessel geometries. A sensitivity study involving the post-processing of outputs with different gas pressures, beam injection angles, magnetic field pitch angles, and system resolutions will benchmark the code against the respective experimental data.

¹Research funded by USDOE DE-AC02-76CH03073.

William Schumaker
Lawrence Technological University

Date submitted: 19 Jul 2007

Electronic form version 1.4