

Abstract Submitted
for the GEC09 Meeting of
The American Physical Society

Gas Temperature Measurement in a Glow Discharge Plasma

KENNETH SLONEKER, Electronic Development Labs Inc., NIRMOL PODDER, WILLIAM E. MCCURDY, SHI SHI, Dept. of Math & Physics, Troy University — In this study a relatively inexpensive quartz protected thermocouple is used to measure the gas temperature in the positive column of a glow discharge plasma. For simplicity a K-type thermocouple is used to interpret the gas temperature from the sensor voltage at pressures from 0.5 Torr to 15 Torr and discharge currents from 5 mA to 120 mA. Gas temperature is investigated as a function of the gas pressure at fixed discharge currents and as a function of discharge current at fixed gas pressures in three different gas species (Ar, N₂, and He). An infinite cylinder model is used to compute the average gas temperature of the discharge from joule heating and gas thermal conductivity. The model and measurement data agree within 1% to 10% depending on plasma parameters. Data for all three gases have a similar quasi-linear increasing error as compared to the model.

Kenneth Sloneker
Electronic Development Labs Inc.

Date submitted: 14 Jul 2009

Electronic form version 1.4