

Abstract Submitted  
for the MAR05 Meeting of  
The American Physical Society

**Detection of NO<sub>2</sub> down to ppb levels using individual and multiple In<sub>2</sub>O<sub>3</sub> nanowire devices** DAIHUA ZHANG, ZUQIN LIU, CHAO LI, CHONGWU ZHOU, University of Southern California — We demonstrate detection of NO<sub>2</sub> down to ppb levels using transistors based on both single and multiple In<sub>2</sub>O<sub>3</sub> nanowires operating at room temperature. This represents orders-of-magnitude improvement over previously reported metal oxide film or nanowire/nanobelt sensors. A comparison between the single and multiple nanowire sensors reveals that the latter have numerous advantages in terms of great reliability, high sensitivity and simplicity in fabrication. Furthermore, selective detection of NO<sub>2</sub> can be readily achieved with multiple-nanowire sensors even with other common chemicals such as NH<sub>3</sub>, O<sub>2</sub>, CO and H<sub>2</sub> around.

Daihua Zhang  
University of Southern California

Date submitted: 01 Dec 2004

Electronic form version 1.4