

Abstract Submitted  
for the MAR07 Meeting of  
The American Physical Society

**Further demonstration of a local evanescent field array coupled (LEAC) biosensor concept** GUANGWEI YUAN, R. POWNALL, M. STEPHENS, D. DANDY, T. CHEN, P. NIKKEL, K. LEAR, ECE, Colorado State University, Fort Collins, CO 80523 — Low-cost, label-free immunoassay biosensors are needed for point-of-care clinical diagnostics, food safety, environmental monitoring, and biosecurity applications. A novel local, evanescent-field, array coupled (LEAC) photonic biosensor that can simultaneously sense multiple viruses, proteins, or DNA oligomers is being investigated. The sensing mechanism relies on the formation of a biological adlayer via specific binding of an analyte target to one of several localized patches of immobilized biological molecule probes (antibodies, ss-DNA, aptamers). The attached analytes modify the waveguide cross-section and thus the optical field. A buried array of evanescently coupled photodetector elements along the length of the waveguide, each opposite a region of specific antibody type, locally sense the modification in the evanescent field due to adlayers of bound analytes. Proof-of-concept experiments have demonstrated strong optical modulation responses to artificial adlayers varying from 17 to  $\sim 100$  nm in thickness as observed by near-field scanning optical microscopy. LEAC sensors with electronic readout circuits are have been fabricated in a commercial  $0.35\ \mu\text{m}$  CMOS technology. Currently, research efforts are characterizing the effect of different polymer and organic molecules adlayers on the detected optical signal using these sensors.

Guangwei Yuan  
ECE, Colorado State University, Fort Collins, CO 80523

Date submitted: 30 Nov 2006

Electronic form version 1.4