

Abstract Submitted
for the MAR15 Meeting of
The American Physical Society

Development and Characterization of Dynamic Light Scattering Instrumentation to Determine Nanoparticle Size¹ T.J. SEBASTIAN, J. HARDING, T. VOLPE, J.R. SIMPSON, M. SCHULZE, S.M. LEV, Towson Univ — Dynamic Light Scattering (DLS) provides a high-throughput and accurate measurement of particle sizes for monodisperse (MD) spherical nanoparticles (NPs). We report on the development and characterization of homebuilt DLS instrumentation to measure the size of MD NPs of gold, polystyrene, and ZnO. HeNe and Ar-ion lasers comprise the excitation sources for the scattering experiment. An avalanche photodiode detects scattered light and an autocorrelation card analyzes the signal to provide a measurement of the translational diffusion coefficient, which for MD and spherical particles allows for the determination of NP radius. We have tested our apparatus using commercially produced gold NPs in the range of 10 nm to 200 nm and synthesized ZnO NPs. DLS measurements were compared to those obtained by Atomic Force Microscopy (AFM). After size characterization, the ZnO NPs will be employed in ongoing toxicity studies.

¹T.J.S. and J.H. acknowledge support from Towson University. J.R.S., M.S. and S.M.L. acknowledge support from NSF - CBET #1236083.

T.J. Sebastian
Towson Univ

Date submitted: 13 Nov 2014

Electronic form version 1.4